



Madhya Pradesh State Electronics Development Corporation Ltd.

State IT Centre, 47-A, Arera Hills, Bhopal – 462 011

CORRIGENDUM-1

Date: 18th January, 2016

Tender No. MPSEDC/ETDC/2015/303 DATED 29.12.2015

Supply and Installation of Calibrators for Testing and Calibration of Equipments as per BIS Standard at MPSEDC Calibration Lab, Bhopal (M.P.)

*The last date of submission of offers has been extended from 20th January, 2016 to **27th January 2016 (Wednesday)** at 3:00 pm. The technical bid will be opened on the same day at 4:00 pm.*

1. 8 ½ DIGIT DIGITAL MULTIMETER PAGE NO. 9

Sr. No.	PARAMETER	SPECIFICATION AS PER TENDER	AMENDED SPECIFICATION
2	Standard function		
C	DC Current	• 8 ranges: 100 nA to 1 A	• 8 ranges: 100 nA to 20 A
		• Up to 1,350 readings/sec (5.5 digits)	• Up to 1,350 readings/sec (5.5 digits)
		• Maximum sensitivity: 1 pA	• Maximum sensitivity: 1 pA
		• 14 ppm 24 hour accuracy	• 14 ppm 24 hour accuracy
D	AC Current	• 5 ranges: 100 µA to 1 A	• 5 ranges: 100 µA to 20 A
		• 10 Hz to 100 kHz bandwidth	• 10 Hz to 100 kHz bandwidth
		• Up to 50 readings/sec	• Up to 50 readings/sec
		• 500 ppm 24 hour accuracy	• 500 ppm 24 hour accuracy

All other terms & conditions remain unchanged.

sd/-

Chief General Manager

M P State Electronics Dev. Corp. Ltd.

State IT Centre, 47-A, Arera Hills Bhopal – 462011,

Phone: 0755-2518754, 2518603 Fax: 0755-2579824