

RESPONSE TO PREBID QUERIES
DATE OF PRE BID MEETING: 07/01/2016

Supply and Installation of Calibrators for Testing and Calibration of Equipments as per BIS Standard at MPSEDC Calibration Lab, Bhopal (M.P.) (Tender No. MPSEDC/ETDC/2015/302 Dated 29/12/2015)

1. ENVIRONMENTAL TEST CHAMBER PAGE NO. 10.

The chamber should be designed for the following tests: Humidity test, Soaking test, Dry heat test, Damp heat test, Cold test, Stress relief test, Ageing processes, Salt spray, Thermal cycle and Preferred climatic categories as per IS specifications.

SI.No.	PARAMETER	ORIGINAL SPECIFICATION AS PER TENDER	RECOMMENDED SPECIFICATION VENDOR: M/S. SOUTHERN SCIENTIFIC LAB INSTRUMENTS CHENNAI THROUGH EMAIL	RESPONSE FROM MPSEDC
4	Humidity Range	10% to 95%, RH, ± 1 to 3% tolerance	40 TO 97% RH	ACCEPTED
13	Ramp Rate	Not mentioned	1°C	ACCEPTED
14	Controls	Not mentioned	Programmable	ACCEPTED
15	Port Holes	Not mentioned	Can be provided	ACCEPTED
16	Specimen Protection	Not Mentioned	Should be provided	ACCEPTED
17	Viewing Window	Not Mentioned	Should be provided	ACCEPTED
SALT SPRAY		The chamber should be designed for the following tests: Humidity test, Soaking test, Dry heat test, Damp heat test, Cold test, Stress relief test, Ageing processes, Salt spray, Thermal cycle and Preferred climatic categories as per IS specifications.	Separate chamber is needed for salt spray test.	ACCEPTED